



HAL
open science

Direct Identification of *Staphylococcus aureus* and Determination of Methicillin Susceptibility From Positive Blood-Culture Bottles in a Bact/ALERT System Using Binax Now . aureus and PBP2a Tests

S. Heraud, A.-M. Freydiere, Anne Doleans-Jordheim, M. Bes, A. Tristan, F. Vandenesch, F. Laurent, O. Dauwalder

► To cite this version:

S. Heraud, A.-M. Freydiere, Anne Doleans-Jordheim, M. Bes, A. Tristan, et al.. Direct Identification of *Staphylococcus aureus* and Determination of Methicillin Susceptibility From Positive Blood-Culture Bottles in a Bact/ALERT System Using Binax Now . aureus and PBP2a Tests. *Annals of Laboratory Medicine*, 2015, 35 (4), pp.454. 10.3343/alm.2015.35.4.454 . hal-02487306

HAL Id: hal-02487306

<https://univ-lyon1.hal.science/hal-02487306v1>

Submitted on 10 Jan 2025

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution - NonCommercial 4.0 International License



Direct Identification of *Staphylococcus aureus* and Determination of Methicillin Susceptibility From Positive Blood-Culture Bottles in a Bact/ALERT System Using Binax Now *S. aureus* and PBP2a Tests

Sandrine Heraud, Pharm.D.¹, Anne-Marie Freydiere, Pharm.D.^{1,2,3,4}, Anne Doleans-Jordheim, Ph.D.^{1,5}, Michèle Bes, Ph.D.^{1,2,3}, Anne Tristan, Ph.D.^{1,2,3,4}, François Vandenesch, M.D.^{1,2,3,4}, Frederic Laurent, Ph.D.^{2,3,4,6,*}, and Olivier Dauwalder, Ph.D.^{1,2,3,4,*}

Hospices Civils de Lyon¹, Laboratoire de Bactériologie, Centre de Biologie et de Pathologie Est, Bron; Hospices Civils de Lyon², Centre National de Référence des Staphylocoques, Centre de Biologie et de Pathologie Est, Bron; International Center for Research in Infectiology³, INSERM U1111, Lyon; Université de Lyon⁴, Faculté de Médecine Lyon Est, Domaine de la Buire, Lyon; Research group on «Bacterial Opportunistic Pathogens and Environment»⁵, UMR 5557 Ecologie Microbienne, CNRS, Université Lyon 1, ENVL, Université de Lyon, Lyon; Hospices Civils de Lyon⁶, Laboratoire de Bactériologie, Centre de Biologie et Pathologie Nord, Lyon, France

Staphylococcus aureus bacteremia is associated with high mortality and morbidity, requiring prompt and appropriate antimicrobial treatment. Therefore, it is important to detect methicillin-resistant *S. aureus* (MRSA) rapidly from blood cultures. Two immunochromatographic tests, BinaxNow *S. aureus* and BinaxNow PBP2a, were directly applied to 79 Bact/Alert bottles that were positive for Gram positive cocci in cluster aggregations. Sensitivity and specificity for the identification of *S. aureus* and determination of methicillin resistance were 94% and 87%, and 100% and 100%, respectively, with less than 30 min of performance time. These tests are efficient and rapid; these tests are valuable alternatives to more sophisticated and expensive methods used in the diagnosis of MRSA bacteremia.

Key Words: *Staphylococcus aureus*, Methicillin resistance, Blood culture, Accuracy, Immunochromatographic test

Staphylococcus aureus bacteremias (SAB), the most common cause of nosocomial bacteremias, are associated with high mortality and morbidity [1]. Methicillin-resistant *S. aureus* (MRSA) bacteremia is characterized by higher mortality than methicillin-susceptible *S. aureus* (MSSA) bacteremia. Thus, as a precaution in the absence of an early, definitive method to detect MRSA, the first-line treatment for nosocomial SAB is vancomycin [1]. However, vancomycin is associated with a higher mortality rate than β -lactams in the treatment of MSSA bacteremia [2]. Thus, rapid

identification of *S. aureus* and determination of methicillin susceptibility are of crucial importance [3]. Conventional diagnosis of SAB requires at least 2-3 days [4]. Mass spectrometric and molecular tools, performed directly on positive blood cultures (BCs), enable diagnoses in less than 4 hr [5-7]. However, these new tools are not yet available in every laboratory.

This study evaluated the accuracy of two immunochromatographic tests (ICT) that may be used directly in BCs: Binax Now *S. aureus* (BNSA) for *S. aureus* identification and Binax Now

Received: September 15, 2014

Revision received: January 6, 2015

Accepted: April 25, 2015

Corresponding author: Olivier Dauwalder
INSERM U1111-Pathogénie Bactérienne
et Immunité Innée-Faculté de Médecine
Lyon Est-Domaine de la Buire 7, rue
Guillaume Paradin, 69008 Lyon, France
Tel: +33-0472-071-839
Fax: +33-0472-071-842
E-mail: olivier.dauwalder@chu-lyon.fr

*Olivier Dauwalder and Frederic Laurent shared in the supervision of this study.

© The Korean Society for Laboratory Medicine

This is an Open Access article distributed under the terms of the Creative Commons Attribution Non-Commercial License (<http://creativecommons.org/licenses/by-nc/3.0>) which permits unrestricted non-commercial use, distribution, and reproduction in any medium, provided the original work is properly cited.

PBP2a (BNPBP2a; Alere SAS, Jouy-en-Josas, France) for determining methicillin resistance. Tests were performed by following the manufacturer's instructions. One hundred colony forming units of 17 MRSA strains, including the 15 primary worldwide MRSA clones, sequence type (ST)1, ST5 (n=2), ST8 (n=2), ST22, ST30, ST45, ST59, ST72, ST80, ST88, ST93, ST228, ST239, ST247, and ST398 clones (French National Reference Center, Lyon, France), were inoculated into 10 mL of fresh human blood from healthy volunteers in charcoal aerobic (FA) and non-charcoal anaerobic (SN) Bact/ALERT BC bottles (bioMérieux, Marcy l'Etoile, France). Following the detection of growth with the 3D Bact/ALERT instrument, a direct examination was performed, and each bottle showing Gram positive cocci in cluster aggregations (GPCCA) was tested with the BNSA test followed by the BNPBP2a test. The BNSA test was positive in 17/17 and 16/17 SN and FA BC bottles, respectively. The single strain that tested negative in FA bottles, an ST45 strain, was positive on retesting. Therefore, the first test was considered as a technical error. The BNPBP2a test was positive in 17/17 and 17/17 of SN and FA BC bottles, respectively.

Next, blood from 60 patients (23 females and 37 males, mean age of 41.6 yr) that were hospitalized in surgical and medical care units of the Hospices Civils de Lyon were prospectively collected in accordance with the ethical board of our institution. To reduce the rate of BC positive to coagulase-negative staphylococci, BC bottles from patients hospitalized (i) for more than 24 hr and (ii) in care units known to have a high rate of BC with coagulase-negative staphylococci (CNS) were excluded. Samples were cultured in 79 bottles and then evaluated from August to December 2010. The inclusion criteria for BCs included a growth-detection time of less than 25 hr and the de-

tection of GPCCA by microscopic examination. ICTs were performed on 38 non-charcoal aerobic (SA) and 41 charcoal anaerobic (FN) bottles within 4 hr of growth detection, and the results were compared with species identification by matrix assisted laser desorption ionization/time of flight mass spectrometry (MALDI-TOF MS) by the Saramis system (bioMérieux, Marcy l'Etoile, France) and with methicillin susceptibility testing by the Phoenix® system (BD, Pont de Claix, France) using subcultures on blood agar plates. Any discrepancy in methicillin susceptibility results was checked by PCR for the presence of the *mecA* gene. Of the 79 BC bottles tested, 73 yielded a mono-microbial culture, and four out of these 73 bottles yielded false positive BNSA results. All four were from charcoal bottles (Table 1). Whereas BNSA testing on Bact/ALERT® bottles was cleared by the FDA [8], this work is the first external study evaluating the combination of BNSA and Bact/ALERT bottles, including some charcoal bottles. As shown in Fig. 1, these four false positive BNSA tests exhibited the expected pink control band and a very low intensity gray color band for the sample. These four bottle samples (three *S. epidemidis* and one *S. capitis*) were retested, and while three showed similar low positive results, one was clearly negative. The gray color of the false positive tests and the lack of similar reports from testing BC bottles without charcoal, may suggest interference with charcoal particles as a cause of the false positives. In contrast, no indeterminate result was obtained with the BNPBP2a test using a filtration protocol instead of centrifugation for the charcoal bottles. Therefore, the filtration protocol seemed to be more appropriate than differential centrifugation, and the manufacturer's protocol should be revised to recommend use of a filtration method or recommending against the use of charcoal-containing media with the BNSA test. Of the

Table 1. Results and performance of the Binax Now *Staphylococcus aureus* tests performed on 79 positive blood cultures with results from direct microscopic examination of Gram positive cocci arranged in cluster aggregations

	Positive BNSA	Negative BNSA
<i>S. aureus</i> blood cultures (N=43)	43	0
CNS blood cultures (N=30)	4	26
<i>S. aureus</i> and CNS mixed blood cultures (N=6)	3	3
Sensitivity (%)	93.9 (83.1-98.7)*	
Specificity (%)	86.7 (69.3-96.2)*	
Positive predictive value (%)	92.0 (80.8-97.8)*	
Negative predictive value (%)	89.7 (72.7-97.8)*	

*(): 95% confidence interval.

Abbreviations: BNSA, Binax Now *Staphylococcus aureus*; CNS, coagulase-negative staphylococci.




Fig. 1. Example of false positive Binax Now *Staphylococcus aureus* test with charcoal particles in Bact/ALERT bottles.

six mixed SA/CNS BC, three tested negative for *S. aureus* using BNSA, suggesting the inoculum was too low. Similar false negative BNSA tests were reported by Dhilman *et al.* [9] and Yossepowitch *et al.* [10], with 2/2 and 4/5 false negative results, respectively, found for mixed SA/CNS BC. False negative results for mixed SA/CNS BC were also observed with molecular testing [6, 11]. Our BNSA results were both less sensitive and less specific than those reported by Qian *et al.* [12] (sensitivity [Se]: 97.6%; specificity [Sp]: 100%) and Dhiman *et al.* [9] (Se: 95.8%; Sp: 99.6%), probably due to their use of charcoal-free BC bottles, Bactec and VersaTREK, respectively.

Consistent with studies by Romero-Gomez *et al.* [4] and Montgomery (21st European Congress of Clinical Microbiology and Infectious Diseases, abstract P1040), BNPBP2a detected 10 MRSA strains among the 43 *S. aureus* BCs (Se: 100%; Sp: 100%; Table 2). Both false positive and false negative results were obtained when detecting methicillin resistance in CNS BCs; however, this test was not designed for CNS strains. Surprisingly, of the six mixed SA/CNS BCs, BNPBP2a testing showed concordant results with the reference method. The sole false positive BNPBP2a test observed with CNS was negative on retesting, suggesting that the first result was an artifact of the procedure. Conversely, false negative results in the detection of methicillin resistance in colonies was previously reported with the ClearView PBP2a ICT and with the Slidex MRSA detection test [13]. To circumvent this lack of sensitivity, two different protocols have been reported: the use of an increased inoculum density or the use of colonies previously induced either by cefoxitin or oxacillin to produce PBP2a [13]. Thus, additional studies are required to determine, if these protocols could be applied to the BNPBP2a test.

Although our study has some limitations due to the limited number of staphylococcal BCs (n=79) and the lack of methicillin resistant strains exhibiting the very rare and newly described *mecC* gene [14], these ICT methods performed well in rapidly detecting *S. aureus*/CNS and MSSA/MRSA. A positive BNSA test result could be transmitted to clinicians as “*S. aureus*,” whereas a negative result must be transmitted only as “negative test,” in association with results of direct examination of blood culture. A negative test could be due to CNS as well as other Gram positive bacteria in clusters (e.g., *Micrococcus* sp.). Similarly, the BNPBP2a results must be limited to a MSSA/MRSA distinction. In comparison, probes using peptide nucleic acid fluorescence *in situ* hybridization technology (PNA-FISH) have enabled the differentiation of *S. aureus* and CNS in less than 15 min, with a sensitivity of 100% and a specificity of 98.5%, but

do not allow for the detection of methicillin resistance [5].

Molecular tests have also been widely evaluated. The Genotype Gram positive test enables the identification of some strains of *Staphylococcus*, and detects *mecA* with accuracy in 5 hr [15]. The GeneOhm assay shows better sensitivity (100%) and specificity (100%) for *S. aureus* detection than BNSA, whereas BNPBP2a shows better results for MRSA detection than GeneOhm, most likely due to the SCC*mec*-positive strains that do not express *mecA* [6]. The fully automated Xpert MRSA/SA BC test is better for *S. aureus* detection (Se: 100%; Sp: 98.6%) but not for the detection of MRSA (Se: 98.3%; Sp: 99.4%) when compared with BNSA and BNPBP2a results, respectively [11]. The MALDI-TOF MS technique can identify *S. aureus* and CNS directly from BC in 94% to 100% and 25% to 100% of cases, respectively, depending on the CNS species, pre-analytical

Table 2. Results and performance of Binax Now PBP2a tests performed on 79 positive blood cultures with results from direct microscopic examination of Gram positive cocci arranged in cluster aggregations

	Positive BNPBP2a	Negative BNPBP2a
MSSA (N=33)	0	33
MRSA (N=10)	10	0
MS-CNS (N=10)	1	9
MR-CNS (N=20)	14	6
MSSA and MS-CNS (N=1)	0	1
MSSA and MR-CNS (N=4)	4	0
MRSA and MR-CNS (N=1)	1	0
Sensitivity on SA (%)	100 (69.2-100)*	
Specificity on SA (%)	100 (89.4-100)*	
Positive predictive value on SA (%)	100 (69.2-100)*	
Negative predictive value on SA (%)	100 (89.4-100)*	
Sensitivity on CNS (%)	70 (45.7-88.1)*	
Specificity on CNS (%)	90.0 (55.5-99.8)*	
Positive predictive value on CNS (%)	95 (68.1-99.8)*	
Negative predictive value on CNS (%)	60 (32.3-83.7)*	
Overall sensitivity (%) [†]	81.3 (63.6-92.8)*	
Overall specificity (%) [†]	97.7 (88.0-99.9)*	
Overall positive predictive value (%) [†]	97 (81.0-99.9)*	
Overall negative predictive value (%) [†]	88 (75.2-93.4)*	

* (): 95% confidence interval; [†]Overall: monomicrobial+polymicrobial cultures.

Abbreviations: BNPBP2a, Binax Now PBP2a; MSSA, methicillin-susceptible *Staphylococcus aureus*; MRSA, methicillin-resistant *Staphylococcus aureus*; MS-CNS, methicillin-susceptible coagulase-negative *Staphylococci*; MR-CNS, methicillin-resistant coagulase-negative *Staphylococci*; SA, *Staphylococcus aureus*; CNS, coagulase-negative *Staphylococci*.

methods, and type of BC bottles that are included in the process [7]. However, thus far, MALDI-TOF MS cannot detect methicillin resistance [16]. An alternative method may be the combination of MALDI-TOF MS for the identification of *S. aureus* and a BNPBP2a test to determine methicillin resistance.

Finally, diagnosis of staphylococcal bacteremia may be hastened with all of these new tools; however, molecular tests are expensive [5] and MALDI-TOF MS, which is less expensive, does not enable methicillin resistance determination [7]. Thus, the combination of BNSA and BNPBP2a tests appear to be an efficient diagnostic strategy, because they are simple, rapid, and feasible for any laboratory personnel.

Authors' Disclosures of Potential Conflicts of Interest

Alere France SAS provided the *Staphylococcus aureus* and the PBP2a Binax Now® tests for this study. Frederic Laurent received a travel grant from Alere France SAS to attend the ECC-MID 2013 meeting. Olivier Dauwalder and Francois Vandenesch received research grants from bioMérieux, outside of the submitted work.

Acknowledgments

We thank Alere France SAS for the kind supply of the *Staphylococcus aureus* and the PBP2a Binax Now® tests for this study. We are also grateful to the lab technicians and laboratory medicine-specialized biologists for their assistance in tracking Gram positive cocci blood cultures.

This study was supported by Hospices Civils de Lyon and Institut de Veille Sanitaire (InVS). Alere France SAS supplied free *Staphylococcus aureus* and PBP2a Binax Now® tests for this study.

REFERENCES

1. de Kraker ME, Wolkewitz M, Davey PG, Koller W, Berger J, Nagler J, et al. Clinical impact of antimicrobial resistance in European hospitals: excess mortality and length of hospital stay related to methicillin-resistant *Staphylococcus aureus* bloodstream infections. *Antimicrob Agents Chemother* 2011;55:1598-605.
2. Kim SH, Kim KH, Kim HB, Kim NJ, Kim EC, Oh MD, et al. Outcome of vancomycin treatment in patients with methicillin-susceptible *Staphylococcus aureus* bacteremia. *Antimicrob Agents Chemother* 2008;52:192-7.
3. Bates DW, Goldman L, Lee TH. Contaminant blood cultures and resource utilization. The true consequences of false-positive results. *JAMA* 1991;265:365-9.
4. Romero-Gómez MP, Quiles-Melero I, Navarro C, Paño-Pardo JR, Gómez-Gil R, Mingorance J. Evaluation of the BinaxNOW PBP2a assay for the direct detection of methicillin resistance in *Staphylococcus aureus* from positive blood culture bottles. *Diagn Microbiol Infect Dis* 2012;72:282-4.
5. Carretto E, Bardaro M, Russello G, Mirra M, Zuelli C, Barbarini D. Comparison of the *Staphylococcus* QuickFISH BC test with the tube coagulase test performed on positive blood cultures for evaluation and application in a clinical routine setting. *J Clin Microbiol* 2013;51:131-5.
6. Gröbner S, Dion M, Plante M, Kempf VA. Evaluation of the BD GeneOhm StaphSR assay for detection of methicillin-resistant and methicillin-susceptible *Staphylococcus aureus* isolates from spiked positive blood culture bottles. *J Clin Microbiol* 2009;47:1689-94.
7. Lagacé-Wiens PR, Adam HJ, Karlowsky JA, Nichol KA, Pang PF, Guenther J, et al. Identification of blood culture isolates directly from positive blood cultures by use of matrix-assisted laser desorption ionization-time of flight mass spectrometry and a commercial extraction system: analysis of performance, cost, and turnaround time. *J Clin Microbiol* 2012;50:3324-8.
8. Alere. BinaxNOW® *Staphylococcus aureus* Card www.accessdata.fda.gov/cdrh_docs/reviews/K090964.pdf (Updated on Dec 16, 2009).
9. Dhiman N, Trienski TL, DiPersio LP, DiPersio JR. Evaluation of the BinaxNOW *Staphylococcus aureus* test for rapid identification of Gram-positive cocci from VersaTREK blood culture bottles. *J Clin Microbiol* 2013;51:2939-42.
10. Yossepowitch O, Dan M, Kutchinsky A, Gottesman T, Schwartz-Harari O. A cost-saving algorithm for rapid diagnosis of *Staphylococcus aureus* and susceptibility to oxacillin directly from positive blood culture bottles by combined testing with BinaxNOW® *S. aureus* and Xpert MRSA/SA Assay. *Diagn Microbiol Infect Dis* 2014;78:352-5.
11. Spencer DH, Sellenriek P, Burnham CA. Validation and implementation of the GeneXpert MRSA/SA blood culture assay in a pediatric setting. *Am J Clin Pathol* 2011;136:690-4.
12. Qian Q, Eichelberger K, Kirby JE. Rapid identification of *Staphylococcus aureus* directly from Bactec blood culture broth by the BinaxNOW *S. aureus* test. *J Clin Microbiol* 2014;52:319-20.
13. Corso A, Soloaga R, Faccione D, Gaggioli P, Corbella S, Iglesias M, et al. Improvement of a latex agglutination test for the evaluation of oxacillin resistance in coagulase-negative staphylococci. *Diagn Microbiol Infect Dis* 2004;50:223-5.
14. Laurent F, Chardon H, Haenni M, Bes M, Reverdy ME, Madec JY, et al. MRSA harboring *mecA* variant gene *mecC*, France. *Emerg Infect Dis* 2012;18:1465-7.
15. Eigner U, Weizenegger M, Fahr AM, Witte W. Evaluation of a rapid direct assay for identification of bacteria and the *mecA* and *van* genes from positive-testing blood cultures. *J Clin Microbiol* 2005;43:5256-62.
16. Szabados F, Kaase M, Anders A, Gatermann SG. Identical MALDI TOF MS-derived peak profiles in a pair of isogenic SCC*mec*-harboring and SCC*mec*-lacking strains of *Staphylococcus aureus*. *J Infect* 2012;65:400-5.